



Heartworm Ag canine ELISA

Enzyme Immunoassay for the detection of heartworm antigen (Dirofilaria Immitis) in serum and plasma samples of canines

VET

REF

DE4795



96 wells

Please use only the valid version of the package insert provided with the kit

Table of Contents

1	INTRODUCTION	3
2	PRINCIPLE OF THE TEST KIT	3
3	CONTENTS	3
4	HANDLING AND STORAGE OF SPECIMENS.....	3
5	WASHING PROTOCOL	4
6	TEST PROTOCOL.....	4
7	PRECAUTIONS	5
8	INTERPRETATION OF TEST RESULTS.....	5
	SYMBOLS USED WITH DEMEDITEC ASSAYS	6

1 INTRODUCTION

For Veterinary use only!

Dirofilaria Immitis (Nematode Filaridae) has been reported parasitizing a variety of wild canids, foxes and felids (9% positive animals found in Australia). Prevalence of infection may vary with geographic location, habitat, densities of mosquito vectors and definitive hosts and climatic conditions. In diseased dogs *Dirofilaria Immitis* are found in heart, lungs, pulmonary arterities or thoracic vena cava. The amount of Microfilaraemia correlates with the number of adult filariae which also correlates with the age and weight of the dog. *Dirofilaria* antigen titers correlates best with weight and worms present ($R=91$) adjusted to female worm equivalents (four male worms equal to one female worm).

Epidemiological studies on the canine population have demonstrated that *Dirofilaria Immitis* is the predominant species prevalence between (0,6% -34%) have been reported (Valladres et al 1984/Perez-Sanches et al 1989) important areas are Italy, France, Spain, Portugal, Australia and Brazil. Different potential vectors have been studied, the ability is attributed to *Aedes Vexans*, *Aedes caspui* and *Aedes longitubes* (blood sucking mosquitoes).

Several studies demonstrate **Human Dirofilariosis** (\pm 250 cases are reported) in high prevalence area the majority of this infections are subclinical. A small percentage of patients with symptoms (Malaise, Thoracic aches, low fever, cough) all have pulmonaire "coin lesions". Some ELISA antigen test also detects cross-reactive *Dirofilaria* SPP (*rapens/Dipetalonema*).

The life cycle of the *Dirofilaria Immitis* is as follows: *Microfilaria* develops (inside mosquito) within 14 days (10-16) are transferred by the mosquito to the host (dog/fox etc.) by biting. This L3 stage in the definitive host moults five times and migrates to the heart area (venous part) within 6 months. Ivermectin and diethylcarbamazine can be used as treatment for infected dogs/cats (0,006mg/kg) once a month.

2 PRINCIPLE OF THE TEST KIT

The heartworm (*Dirofilaria Immitis*) ELISA antigen test kit detects heat stable *Dirofilaria Immitis* antigens, which is the major course of canine Heartworm in serum samples.

The principle of the test is based on the reaction of two monoclonal antibodies with an antigenic determinant against *Dirofilaria Immitis*. One monoclonal antibody, coated to the plate, catches the *Dirofilaria* in the serum or plasma sample after which the other, enzyme-labeled antibody detects the bound antigen.

After incubation and rinsing, the substrate is then added and the optical density is measured at 450 nm.

3 CONTENTS

- 12 x 8-well **microtiter strips** coated with monoclonal anti-*Dirofilaria* antibody.
- 1 x **strip holder**
- 1 x 11 ml **HRPO-conjugated** monoclonal **antibody**
- 1 x 1 ml **positive control** (Ready to use)
- 1 x 1 ml **negative control** (Ready to use)
- 1 x 20 ml **wash solution** 200 x concentrated (*must be diluted in deionized water before use!*)
- 1 x 15 ml **ELISA buffer**
- 1 x 8 ml **substrate buffer A**
- 1 x 8 ml **substrate buffer B**
- 1 x 8 ml **stop solution**
- 1 x plastic **cover seal**

4 HANDLING AND STORAGE OF SPECIMENS

The kit should be stored at +4°C. An open packet should be used within 10 days.

Samples may be used fresh or may be kept frozen below -20°C before use.

Positive and negative controls may be stored after reconstitution in aliquots at -20°C and used until the expiry date. Avoid repeated freezing and thawing as this increases non-specific reactivity.

5 WASHING PROTOCOL

In ELISA's un-complexed components must be removed efficiently between each immunological incubation step. This is accomplished by appropriate washing. It should be stressed that each washing step must be carried out with care to guarantee reproducible inter- and intra assay results. It is advised to follow the washing procedures outlined below carefully.

Both manual washing and washing with automatic equipment can be performed. (Automatic washing equipment usually gives better results).

Manual washing

1. Empty each well by turning the microtiter plate upside down followed by a firm vertical downward movement to remove the buffer.
2. Fill all the wells with 250 μ l washing solution.
3. This washing cycle (1 and 2) should be carried out at least 4 times.
4. Turn the plate upside down and empty the wells by a firm short vertical movement.
5. Place the inverted plate on absorbent paper towels and tap the plate firmly to remove any residual washing solution from the wells.
6. Take care that none of the wells dries out before the next reagent is dispensed.

Washing with automatic equipment

When using automatic plate washing equipment, check that all wells can be aspirated completely and that the washing solution is correctly dispensed, reaching the rim of each well during each rinsing cycle. The washer should be programmed to execute at least 4 washing cycles.

6 TEST PROTOCOL

1. Open the packet of strips, take out the strips to be used, cover the remaining strips with a part of the provided seal and store them at +4 °C and use them within 10 days.
 2. Wash the microtiter strip(s) with washing solution, according to the washing protocol.
 3. Add 100 μ L positive control to one well.
Add 100 μ L negative control to the next well.
 4. Add 50 μ L buffer to all remaining wells and thereafter add 50 μ L serum of each serum sample to be tested to the subsequent well.
 5. Seal and incubate for 60 min. at 37 °C.
 6. Wash as in 2.
 7. Dispense 100 μ L anti-Heartworm antibody conjugate to all wells.
 8. Seal and incubate for 60 min. at 37 °C.
 9. Wash as in 2.
 10. Mix equal parts of buffer A and buffer B together with gentle shaking. Prepare immediately before use!
Dispense 100 μ L substrate solution to each well.
Incubate 10-15 min. at room temperature.
 11. Add 50 μ L stop solution to each well.
 12. Read the absorbency values immediately (within 10 min.!) at 450 nm ref. 620 nm
-

7 PRECAUTIONS

- Handle all biological materials as though capable of transmitting infectious diseases.
- Do not pipette by mouth.
- Do not eat, drink, smoke, prepare foods, or apply cosmetics within the designated work area
- TMB is toxic by inhalation, in contact with skin and when swallowed; observe care when handling the substrate.
- Do not use components past expiration date and do not intermix components from different serial lots.
- Optimal results will be obtained by strict adherence to this protocol. Careful pipetting and washing throughout this procedure are necessary to maintain precision and accuracy.
- Each well is ultimately used as an optical cuvette. Therefore, do not touch the under-surface of the microtitre plate and prevent it from damage and dirt.

8 INTERPRETATION OF TEST RESULTS

A sample is considered positive when the measured extinction is higher than the OD of the negative control, but at least $\geq 0,300$.











All signals below 0,300 are considered to be negative.











The OD of the positive control must be at least > 0.800 .

In case of a negative result, the presence of *Microfilaria* should be tested by microscope.

The entire risk as to the performance of these products is assumed by the purchaser. DEMEDITEC shall not be liable for indirect, special or consequential damages of any kind resulting from use of the products.

SYMBOLS USED WITH DEMEDITEC ASSAYS

Symbol	English	Deutsch	Français	Español	Italiano
	Consult instructions for use	Gebrauchsanweisung beachten	Consulter les instructions d'utilisation	Consulte las instrucciones de uso	Consultare le istruzioni per l'uso
	European Conformity	CE-Konformitätskennzeichnung	Conformité aux normes européennes	Conformidad europea	Conformità europea
	In vitro diagnostic device	In-vitro-Diagnostikum	Usage Diagnostic in vitro	Para uso Diagnóstico in vitro	Per uso Diagnostica in vitro
	For research use only	Nur für Forschungszwecke	Seulement dans le cadre de recherches	Sólo para uso en investigación	Solo a scopo di ricerca
	Catalogue number	Katalog-Nr.	Numéro de catalogue	Número de catálogo	Numero di Catalogo
	Lot. No. / Batch code	Chargen-Nr.	Numéro de lot	Número de lote	Numero di lotto
	Contains sufficient for <n> tests/	Ausreichend für "n" Ansätze	Contenu suffisant pour "n" tests	Contenido suficiente para <n> ensayos	Contenuto sufficiente per "n" saggi
	Storage Temperature	Lagerungstemperatur	Température de conservation	Temperatura de conservación	Temperatura di conservazione
	Expiration Date	Mindesthaltbarkeitsdatum	Date limite d'utilisation	Fecha de caducidad	Data di scadenza
	Legal Manufacturer	Hersteller	Fabricant	Fabricante	Fabbricante
Distributed by	Distributor	Vertreiber	Distributeur	Distribuidor	Distributore
Content	Content	Inhalt	Conditionnement	Contenido	Contenuto
Volume/No.	Volume / No.	Volumen/Anzahl	Volume/Quantité	Volumen/Número	Volume/Quantità

Symbol	Portugues	Dansk	Svenska	Ελληνικά
	Consulte as instruções de utilização	Se brugsanvisning	Se bruksanvisningen	Εγχειρίδιο χρήστη
	Conformidade com as normas europeias	Europaeisk overensstemmelse	Europeisk överensstämmelse	Ευρωπαϊκή Συμμόρφωση
	Diagnóstico in vitro	In vitro diagnostik	Diagnostik in vitro	in vitro διαγνωστικό
				
	Catálogo n.º	Katalognummer	Katalog nummer	Αριθμός καταλόγου
	No do lote	Lot nummer	Batch-nummer	Αριθμός Παρτίδος
		Indeholder tilstrækkeligt til "n" test	Innehåller tillräckligt till "n" tester	Περιεχόμενο επαρκές για «n» εξετάσεις
	Temperatura de conservação de	Opbevarings-temperatur	Förvaringstemperatur	Θερμοκρασία αποθήκευσης
	Prazo de validade	Udløbsdato	Bäst före datum	Ημερομηνία λήξης
	Fabricante	Producent	Tillverkare	Κατασκευαστής
Distributed by				
Content	Conteúdo	Indhold	Innehåll	Περιεχόμενο
Volume/No.	Volume/Número	Volumen/antal	Volym/antal	Όγκος/αριθ..