


RayBio® Mouse Galectin-7 ELISA Kit

**User Manual
(Revised Mar 1, 2012)**

**RayBio® Mouse Galectin-7
ELISA Kit Protocol**

(Cat#: ELM-Galectin7-001)



RayBiotech, Inc.

**We Provide You With Excellent
Protein Array System And Service**

**Tel:(Toll Free)1-888-494-8555 or 770-729-2992; Fax:770-206-2393;
Web: www.raybiotech.com Email: info@raybiotech.com**



RayBiotech, Inc.

**RayBio® Mouse Galectin-7
ELISA Kit Protocol**

TABLE OF CONTENTS

| | | |
|-------|------------------------------------|----|
| I. | Introduction..... | 2 |
| II. | Reagents..... | 2 |
| III. | Storage..... | 3 |
| IV. | Additional Materials Required..... | 3 |
| V. | Reagent Preparation..... | 3 |
| VI. | Assay Procedure..... | 5 |
| VII. | Assay Procedure Summary..... | 6 |
| VIII. | Calculation of Results | |
| A. | Typical Data..... | 7 |
| B. | Sensitivity..... | 8 |
| C. | Recovery..... | 9 |
| D. | Linearity..... | 9 |
| E. | Reproducibility..... | 9 |
| IX. | Specificity..... | 9 |
| X. | Troubleshooting Guide..... | 10 |

I. INTRODUCTION

The RayBio® Mouse Galectin-7 ELISA (Enzyme-Linked Immunosorbent Assay) kit is an *in vitro* enzyme-linked immunosorbent assay for the quantitative measurement of Mouse Galectin-7 in serum, plasma (recommended using heparin and citrate as an anticoagulant), and cell culture supernatants. This assay employs an antibody specific for Galectin-7 coated on a 96-well plate. Standards and samples are pipetted into the wells and Galectin-7 present in a sample is bound to the wells by the immobilized antibody. The wells are washed and biotinylated anti-Mouse Galectin-7 antibody is added. After washing away unbound biotinylated antibody, HRP-conjugated streptavidin is pipetted to the wells. The wells are again washed, a TMB substrate solution is added to the wells and color develops in proportion to the amount of Galectin-7 bound. The Stop Solution changes the color from blue to yellow, and the intensity of the color is measured at 450 nm.

II. REAGENTS

1. Galectin-7 Microplate (Item A): 96 wells (12 strips x 8 wells) coated with anti-Mouse Galectin-7.
2. Wash Buffer Concentrate (20x) (Item B): 25 ml of 20x concentrated solution.
3. Standards (Item C): 2 vials, recombinant Mouse Galectin-7.
4. Assay Diluent C (Item L): 30 ml of diluent buffer. For Standard/ Sample (serum/plasma samples/cell culture supernatants) diluent.
5. Assay Diluent B (Item E): 15 ml of 5x concentrated buffer. For detection antibody and HRP-Streptavidin diluent.
6. Detection Antibody Galectin-7 (Item F): 2 vial of biotinylated anti-Mouse Galectin-7 (each vial is enough to assay half microplate).
7. HRP-Streptavidin concentrate (Item G): 200 μ l 160x concentrated HRP-conjugated streptavidin.
8. TMB One-Step Substrate Reagent (Item H): 12 ml of 3,3',5,5'-tetramethylbenzidine (TMB) in buffered solution.

9. Stop Solution (Item I): 8 ml of 0.2 M sulfuric acid.

III. STORAGE

May be stored for up to 6 months at 2° to 8°C from the date of shipment. Standard (recombinant protein) should be stored at -20°C or -80°C (recommended at -80°C) after reconstitution. Opened Microplate Wells or reagents may be stored for up to 1 month at 2° to 8°C. Return unused wells to the pouch containing desiccant pack, reseal along entire edge. Note: the kit can be used within one year if the whole kit is stored at -20°C. Avoid repeated freeze-thaw cycles.

IV. ADDITIONAL MATERIALS REQUIRED

- 1 Microplate reader capable of measuring absorbance at 450 nm.
- 2 Precision pipettes to deliver 2 µl to 1 ml volumes.
- 3 Adjustable 1-25 ml pipettes for reagent preparation.
- 4 100 ml and 1 liter graduated cylinders.
- 5 Absorbent paper.
- 6 Distilled or deionized water.
- 7 Log-log graph paper or computer and software for ELISA data analysis.
- 8 Tubes to prepare standard or sample dilutions.


V. REAGENT PREPARATION

1. Bring all reagents and samples to room temperature (18 - 25°C) before use.
2. Sample dilution: If your samples need to be diluted, Assay Diluent C (Item L) should be used for dilution of serum/plasma/culture supernatants.

Suggested dilution for normal serum/plasma: 2 fold*.

* Please note that levels of the target protein may vary between different specimens. Optimal dilution factors for each sample must be determined by the investigator.

3. Assay Diluent B (Item E) should be diluted 5-fold with deionized or distilled water before use.
4. Preparation of standard: **Briefly spin the vial of Item C.** Add 400 μl Assay Diluent C (Item L) into Item C vial to prepare a 50 ng/ml standard solution. **Dissolve the powder thoroughly by a gentle mix.** Add 180 μl 50 ng/ml standard from the vial of Item C, into a tube with 320 μl Assay Diluent C to prepare a 16,000 pg/ml standard solution. Pipette 300 μl Assay Diluent C into each tube. Use the 16,000 pg/ml standard solution to produce a dilution series (shown below). Mix each tube thoroughly before the next transfer. Assay Diluent C serves as the zero standard (0 pg/ml).



5. If the Wash Concentrate (20x) (Item B) contains visible crystals, warm to room temperature and mix gently until dissolved. Dilute 20

ml of Wash Buffer Concentrate into deionized or distilled water to yield 400 ml of 1x Wash Buffer.

6. Briefly spin the Detection Antibody vial (Item F) before use. Add 100 μ l of 1x Assay Diluent B (Item E) into the vial to prepare a detection antibody concentrate. Pipette up and down to mix gently (the concentrate can be stored at 4°C for 5 days). The detection antibody concentrate should be diluted 80-fold with 1x Assay Diluent B and used in step 4 of Part VI Assay Procedure.
7. Briefly spin the HRP-Streptavidin concentrate vial (Item G) and pipette up and down to mix gently before use. HRP-Streptavidin concentrate should be diluted 160-fold with 1x Assay Diluent B (Item E).

For example: Briefly spin the vial (Item G) and pipette up and down to mix gently . Add 50 μ l of HRP-Streptavidin concentrate into a tube with 8 ml 1x Assay Diluent B to prepare a final 160 fold diluted HRP-Streptavidin solution. Mix well.

VI. ASSAY PROCEDURE:

1. Bring all reagents and samples to room temperature (18 - 25°C) before use. It is recommended that all standards and samples be run at least in duplicate.
2. Add 100 μ l of each standard (see Reagent Preparation step 2) and sample into appropriate wells. Cover well and incubate for 2.5 hours at room temperature or over night at 4°C with gentle shaking.
3. Discard the solution and wash 4 times with 1x Wash Solution. Wash by filling each well with Wash Buffer (300 μ l) using a multi-channel Pipette or autowasher. Complete removal of liquid at each step is

essential to good performance. After the last wash, remove any remaining Wash Buffer by aspirating or decanting. Invert the plate and blot it against clean paper towels.

4. Add 100 μ l of 1x prepared biotinylated antibody (Reagent Preparation step 6) to each well. Incubate for 1 hour at room temperature with gentle shaking.
5. Discard the solution. Repeat the wash as in step 3.
6. Add 100 μ l of prepared Streptavidin solution (see Reagent Preparation step 7) to each well. Incubate for 45 minutes at room temperature with gentle shaking.
7. Discard the solution. Repeat the wash as in step 3.
8. Add 100 μ l of TMB One-Step Substrate Reagent (Item H) to each well. Incubate for 30 minutes at room temperature in the dark with gentle shaking.
9. Add 50 μ l of Stop Solution (Item I) to each well. Read at 450 nm immediately.

VII. ASSAY PROCEDURE SUMMARY

1. Prepare all reagents, samples and standards as instructed.



2. Add 100 μ l standard or sample to each well.
Incubate 2.5 hours at room temperature or over night at 4°C.



3. Add 100 μ l prepared biotin antibody to each well.
Incubate 1 hour at room temperature.



4. Add 100 μ l prepared Streptavidin solution.
Incubate 45 minutes at room temperature.



5. Add 100 μ l TMB One-Step Substrate Reagent to each well.
Incubate 30 minutes at room temperature.




6. Add 50 μ l Stop Solution to each well.
Read at 450 nm immediately.

VIII. CALCULATION OF RESULTS

Calculate the mean absorbance for each set of duplicate standards, controls and samples, and subtract the average zero standard optical density. Plot the standard curve on log-log graph paper or using Sigma plot software, with standard concentration on the x-axis and absorbance on the y-axis. Draw the best-fit straight line through the standard points.

A. TYPICAL DATA

These standard curves are for demonstration only. A standard curve must be run with each assay.



B. SENSITIVITY

The minimum detectable dose of Galectin-7 is typically less than 65 pg/ml.

C. RECOVERY

Recovery was determined by spiking various levels of Galectin-7 into normal Mouse serum, plasma (heparin and citrate) and cell culture media. Mean recoveries are as follows:

| Sample Type | Average % Recovery | Range (%) |
|--------------------|--------------------|-----------|
| Serum | 141.4 | 133-149 |
| Plasma | 107.3 | 77-135 |
| Cell culture media | 139.0 | 118-148 |

D. LINEARITY

| Sample Type | | Serum | Plasma | Cell Culture Media |
|-------------|-----------------------|-------|--------|--------------------|
| 1:2 | Average % of Expected | 87.25 | 96.94 | 108.5 |
| | Range (%) | 80-94 | 74-120 | 101-116 |
| 1:4 | Average % of Expected | 75.88 | 76.95 | 75.79 |
| | Range (%) | 67-82 | 69-86 | 67-83 |

E. REPRODUCIBILITY

Intra-Assay: CV<10%

Inter-Assay: CV<12%

IX. SPECIFICITY

Cross Reactivity: This ELISA kit shows no cross-reactivity with the following cytokines tested: Mouse CD30, L CD30, T CD40, CRG-2, CTACK, Eotaxin, Eotaxin-2, Fas Ligand, Fractalkine, GCSF, GM-CSF, IFN- γ , IGFBP-3, IGFBP-5, IGFBP-6, IL-1 α , IL-1 β , IL-2, IL-3, IL-3 Rb, IL-4, IL-5, IL-6, IL-9, IL-10, IL-12 p40/p70, IL-12 p70, IL-13, IL-17, KC, Leptin R, LEPTIN(OB), LIX, L-Selectin, Lymphotactin, MCP-1, MCP-5, M-CSF, MIG, MIP-1 α , MIP-1 γ , MIP-2, MIP-3 β , MIP-3 α , PF-4, P-Selectin, RANTES, SCF, SDF-1 α , TARC, TCA-3, TECK, TIMP-1, TNF- α , TNF RI, TNF RII, TPO, VCAM-1, VEGF.

X. TROUBLESHOOTING GUIDE


| Problem | Cause | Solution |
|------------------------|---|---|
| 1. Poor standard curve | <ol style="list-style-type: none"> 1. Inaccurate pipetting 2. Improper standard dilution | <ol style="list-style-type: none"> 1. Check pipettes 2. Ensure briefly spin the vial of Item C and dissolve the powder thoroughly by a gentle mix. |
| 2. Low signal | <ol style="list-style-type: none"> 1. Too brief incubation times 2. Inadequate reagent volumes or improper dilution | <ol style="list-style-type: none"> 1. Ensure sufficient incubation time; assay procedure step 2 change to over night 2. Check pipettes and ensure correct preparation |
| 3. Large CV | <ol style="list-style-type: none"> 1. Inaccurate pipetting | <ol style="list-style-type: none"> 1. Check pipettes |
| 4. High background | <ol style="list-style-type: none"> 1. Plate is insufficiently washed 2. Contaminated wash buffer | <ol style="list-style-type: none"> 1. Review the manual for proper wash. If using an a plate washer, check that all ports are unobstructed. 2. Make fresh wash buffer |
| 5. Low sensitivity | <ol style="list-style-type: none"> 1. Improper storage of the ELISA kit 2. Stop solution | <ol style="list-style-type: none"> 1. Store your standard at $<-20^{\circ}\text{C}$ after reconstitution, others at 4°C. Keep substrate solution protected from light 2. Stop solution should be added to each well before measure |

RayBio® ELISA kits:

Over 200 ELISA kits, custom ELISA kit choose from over 500 list visit www.raybiotech.com for details.

Note:

This product is for research use only.



©2004 RayBiotech, Inc.