

## Lipase Activity Assay Kit III

(Catalog #K724-100; 100 assays; Store kit at -20°C)

**I. Introduction:**

Lipase is a subclass of the esterases that catalyze the hydrolysis of ester bonds of lipid substrates. Lipases perform essential roles in the digestion, transport and processing of dietary lipids (e.g. triglycerides, fats, oils) in most, if not all, living organisms. In humans, pancreatic lipase is the key enzyme responsible for breaking down fats in the digestive system by converting triglycerides to monoglycerides and free fatty acids. During the damage of the pancreas, lipase levels can rise 5- to 10-fold within 24 to 48 hours. In BioVision's Lipase Assay Kit III, Lipase hydrolyzes a specific substrate to generate the methylresorufin, which can be detected fluorometrically at Em/Ex = 529/600 nm. The kit provides a rapid, simple, more sensitive, and reliable test suitable for high throughput assay of Lipase activity. This kit can be used to detect Lipase as low as 0.1 mU/well.

**II. Kit Contents:**

Components	100 assays	Cap Code	Part Number
Lipase Assay Buffer	25 ml	WM	K724-100-1
Lipase Substrate	200 µl	Red	K724-100-2
Methylresorufin Standard (0.1 mM)	40 µl	Yellow	K724-100-3
Lipase Positive Control (lyophilized)	1 vial	Purple	K724-100-4

**III. Storage and Handling:**

Store the kit at -20°C, protect from light. Allow Assay Buffer to warm to room temperature before use. Briefly centrifuge vials before opening. Read the entire protocol before performing the assay.

**IV. Reagent preparation:**

**Lipase Positive Control:** Reconstitute with 100 µl assay buffer. Mix 2 µl Positive Control with 998 µl Lipase Assay Buffer; add 2 µl of the diluted Positive Control into a well and adjust the volume to 50 µl/well with Lipase Assay Buffer. Discard the remaining diluted Positive Control after each use. Aliquot and store the reconstituted Positive Control solution at -20°C. Use within two months.

**V. Lipase Assay Protocol:**

**1. Standard Curve Preparation:**

Add 10 µl of the 0.1mM Methylresorufin Standard to 90 µl Lipase Assay Buffer to generate a 10 µM standard solution. Add 0, 2, 4, 6, 8, 10 µl to each well individually. Adjust the volume to 100 µl/well with Lipase Assay Buffer to generate 0, 20, 40, 60, 80, 100 pmol/well of Methylresorufin Standard. Read fluorometrically at Ex/Em = 529/600nm.

**2. Sample Preparations:**

Tissues (50 mg) or cells (1×10<sup>6</sup>) can be homogenized in ~ 200 µl ice-cold Lipase Assay Buffer then centrifuged to remove insoluble material at 13,000 x g, 10 min. Serum sample can be directly diluted in the Lipase Assay Buffer. Prepare test samples of up to 50 µl/well with Lipase Assay Buffer in a 96-well plate. We suggest testing several doses of your sample to make sure the readings are within the standard curve range.

**3. Reaction Mix:** Mix enough reagents for the number of assays to be performed. For each well, prepare a total 50 µl Reaction Mix.

- 48 µl Assay Buffer
- 2 µl Lipase substrate

Add 50 µl of the Reaction Mixes to each well containing the samples and positive controls. Mix well. Include a reagent background control by adding 50 µl assay buffer to 50 µl reaction mix into a well.

**4. Measurement:** Read Ex/Em = 529/600nm R<sub>1</sub> for sample and R<sub>1B</sub> for background control at T<sub>1</sub>. Read R<sub>2</sub> for sample and R<sub>2B</sub> for background control again at T<sub>2</sub> after incubating the reaction at 37°C for 30 - 60 min (or incubate longer time if the Lipase activity is low), protect from light. The fluorescence generated by the hydrolysis of the Lipase substrate is ΔRFU = (R<sub>2</sub> - R<sub>2B</sub>) - (R<sub>1</sub> - R<sub>1B</sub>). It is recommended to read the fluorescence kinetically to choose the R<sub>1</sub> and R<sub>2</sub> within the linear range of the standard curve.

**5. Calculation:** Subtract the 0 Standard from all Standard readings. Plot the Standard Curve. Apply the ΔRFU to the standard curve to get B nmol of methylresorufin (amount of methylresorufin generated between T<sub>1</sub> and T<sub>2</sub> in the reaction wells):

$$\text{Lipase Activity} = \frac{B}{(T_2 - T_1) \times V} \times \text{Sample Dilution Factor} = \text{nmol/min/ml} = \text{mU/ml}$$


Where: B is the methylresorufin amount from Standard Curve (in nmol).

T<sub>1</sub> is the time of the first reading (R<sub>1</sub>) (in min).

T<sub>2</sub> is the time of the second reading (R<sub>2</sub>) (in min).

V is the pretreated sample volume added into the reaction well (in ml).

**Unit Definition:** One unit is defined as the amount of enzyme that hydrolyzes the substrate to yield 1.0 µmol of methylresorufin per minute at 37°C.



**RELATED PRODUCTS:**

- Lipase Assay Kit I and II
- Glucose Assay Kit
- Pyruvate Assay Kit
- Creatine Assay Kit
- Triglyceride Assay Kit
- Fatty Acid Assay Kit
- Uric Acid Assay Kit
- Lactate Assay Kit I/II
- Creatinine Assay Kit
- Free Glycerol Assay Kit

**FOR RESEARCH USE ONLY! Not to be used on humans.**

”

**GENERAL TROUBLESHOOTING GUIDE:**

Problems	Cause	Solution
Assay not working	<ul style="list-style-type: none"> <li>• Use of ice-cold assay buffer</li> <li>• Omission of a step in the protocol</li> <li>• Plate read at incorrect wavelength</li> <li>• Use of a different 96-well plate</li> </ul>	<ul style="list-style-type: none"> <li>• Assay buffer must be at room temperature</li> <li>• Refer and follow the data sheet precisely</li> <li>• Check the wavelength in the data sheet and the filter settings of the instrument</li> <li>• Fluorescence: Black plates (clear bottoms) ; Luminescence: White plates ; Colorimeters: Clear plates</li> </ul>
Samples with erratic readings	<ul style="list-style-type: none"> <li>• Use of an incompatible sample type</li> <li>• Samples prepared in a different buffer</li> <li>• Cell/ tissue samples were not completely homogenized</li> <li>• Samples used after multiple free-thaw cycles</li> <li>• Presence of interfering substance in the sample</li> <li>• Use of old or inappropriately stored samples</li> </ul>	<ul style="list-style-type: none"> <li>• Refer data sheet for details about incompatible samples</li> <li>• Use the assay buffer provided in the kit or refer data sheet for instructions</li> <li>• Use Dounce homogenizer (increase the number of strokes); observe for lysis under microscope</li> <li>• Aliquot and freeze samples if needed to use multiple times</li> <li>• Troubleshoot if needed</li> <li>• Use fresh samples or store at correct temperatures until use</li> </ul>
Lower/ Higher readings in Samples and Standards	<ul style="list-style-type: none"> <li>• Improperly thawed components</li> <li>• Use of expired kit or improperly stored reagents</li> <li>• Allowing the reagents to sit for extended times on ice</li> <li>• Incorrect incubation times or temperatures</li> <li>• Incorrect volumes used</li> </ul>	<ul style="list-style-type: none"> <li>• Thaw all components completely and mix gently before use</li> <li>• Always check the expiry date and store the components appropriately</li> <li>• Always thaw and prepare fresh reaction mix before use</li> <li>• Refer datasheet &amp; verify correct incubation times and temperatures</li> <li>• Use calibrated pipettes and aliquot correctly</li> </ul>
Readings do not follow a linear pattern for Standard curve	<ul style="list-style-type: none"> <li>• Use of partially thawed components</li> <li>• Pipetting errors in the standard</li> <li>• Pipetting errors in the reaction mix</li> <li>• Air bubbles formed in well</li> <li>• Standard stock is at an incorrect concentration</li> <li>• Calculation errors</li> <li>• Substituting reagents from older kits/ lots</li> </ul>	<ul style="list-style-type: none"> <li>• Thaw and resuspend all components before preparing the reaction mix</li> <li>• Avoid pipetting small volumes</li> <li>• Prepare a master reaction mix whenever possible</li> <li>• Pipette gently against the wall of the tubes</li> <li>• Always refer the dilutions in the data sheet</li> <li>• Recheck calculations after referring the data sheet</li> <li>• Use fresh components from the same kit</li> </ul>
Unanticipated results	<ul style="list-style-type: none"> <li>• Measured at incorrect wavelength</li> <li>• Samples contain interfering substances</li> <li>• Use of incompatible sample type</li> <li>• Sample readings above/below the linear range</li> </ul>	<ul style="list-style-type: none"> <li>• Check the equipment and the filter setting</li> <li>• Troubleshoot if it interferes with the kit</li> <li>• Refer data sheet to check if sample is compatible with the kit or optimization is needed</li> <li>• Concentrate/ Dilute sample so as to be in the linear range</li> </ul>
<p><b>Note:</b> The most probable list of causes is under each problem section. Causes/ Solutions may overlap with other problems.</p>		